



America

CERTIFICATE

No. U8 09 03 66823 002

Holder of Certificate: Cell Science Systems Limited Corp.

852 South Military Trail
Deerfield Beach FL 33442
USA

Production Facility(ies): 66823

Certification Mark:



Product: Measuring instruments for electrical dimensions
Automated Liquid Handling System

Model(s): ROBOCat II

Parameters:

Rated Input Voltage:	220/110 V AC
Rated Input Current:	2 A
Rated Input Power:	200W
Rated Frequency:	50-60 Hz

Tested according to: CAN/CSA C22.2 No. 61010-1:2004
UL 61010-1:2004
EN 61010-2-101:2002
EN 61010-1:2001

The product was voluntarily tested according to the relevant safety requirements and mentioned properties. It can be marked with the certification mark shown above. The certification mark must not be altered in any way. See also notes overleaf.

Test report no.: 090-902174-000

Date, 2009-03-24

Page 1 of 1

